



EXAMPLE REPORT

Date:
January 1, 2021

EXAMPLE
ADDRESS
ADDRESS

Your Address

Dear Clean Water Lab Client:

Sample submission date

Enclosed is the results report for the sample(s) received **Friday, January 01, 2021**

The following results pertain only to the samples listed in the report. The methods used to determine the results have been verified for quality assurance purposes. At the end of the report you can find a summary of our quality control parameters. Please note that the quality control report demonstrates that our testing procedures met a certain quality standard before your samples were analyzed. You can visit our website for more information about our testing analytes.

If you have any questions, please contact me.

Sincerely,

Liz Stacks
Laboratory Director
Clean Water Lab
liz.stacks@cleanwaterlabtech.com

Invoice # 1234 Paid

LABORATORY REPORT

This report shall only be reproduced in full with written consent from Clean Water Lab, Inc.



Report of Analysis

SAMPLE RESULTS SECTION

Lab ID	Sample ID	Analyte	Result	Units	Analysis Start Date	Time
123-TS-1	WELL	E. coli Qualitative	Present	Pres/Abs	1/1/2021	1:00 AM
		Coliform Qualitative	Absent	Pres/Abs	1/2/2021	2:00 AM
		Hardness		40 mg/L CaCO3	1/3/2021	3:00 AM
		pH		1 pH units	1/4/2021	4:00 AM
		Nitrates		5 NO3-N	1/5/2021	5:00 AM
		Nitrites		0 NO2-N	1/6/2021	6:00 AM

Report Author EXAMP
Date 1/1/2021

END OF RESULTS REPORTING SECTION

Sample Information

CHAIN OF CUSTODY INFORMATION

Lab ID	Sample ID	Sample Address	Sample Collector	Sample Collection Date	Time
123-TS-1	WELL	Address Adress	Example	1/1/2021	1:00 AM

Submission

Date 1/1/2021 Time 1:00 AM Submitted By EXAMPLE

LABORATORY REPORT

This report shall only be reproduced in full with written consent from Clean Water Lab, Inc.



Certificates

CERTIFICATES SECTION: NOT RESULTS

Certifying Body	Effective Until	Lab Code	Method Code	Method Description
NC Department of Health and Human Services	July 31, 2021	37900	9223 B	Colilert
NC Department of Health and Human Services	July 31, 2021	37900	Colisure	Colisure
NC Department of Health and Human Services	July 31, 2021	37900	9223B	Colilert Multiple Well
NC Department of Health and Human Services	July 31, 2021	37900	Colisure	Colisure
NC Department of Health and Human Services	July 31, 2021	37900	9223B-MW	Colilert

Quality Control Report QUALITY REPORTS: NOT YOUR RESULTS

This is not the results of sample analysis. This section demonstrates that the laboratory reagents and/or instruments were performing to a certain standard at the time of testing.

Analyte	Method	Accept Reject	QC Sample Range		QC Sample Reading		Date
			Value	Units	Value	Units	
Iron	Method 8008	Accept	1.0 ± 0.15	mg/L Fe	1	mg/L Fe	1/1/2020 ✓
Nitrates	Method 10206	Accept	5.0 ± 0.75	mg/L NO ₃ -N	5	mg/L NO ₃ -N	1/2/2020 ✓
Nitrites	Method 10237	Accept	3.0 ± 0.45	mg/L NO ₂ -N	3.1	mg/L NO ₂ -N	1/3/2020 ✓
Hardness(carbonate)	Method 2340 C	Accept	40.0 ± 6.0	mg/L CaCO ₃	40.1	mg/L CaCO ₃	1/4/2020 ✓
pH	Method 150.3	Accept	5.00 ± 0.01	pH units	5.01	pH units	1/5/2020 ✓

Quality control acceptable range

Analyte	Method	Accept Reject	Reagents Testing Results		Analysis Start	
			EC	TC	Date	Time
E. coli	Colilert	Accept	+	+	8/20/2020	1:59 PM ✓
Coliform	Colilert	Accept	-	+	8/20/2020	1:59 PM ✓
Black Sample	Colilert	Accept	-	-	Final pH 7.26	8/20/2020 1:59 PM ✓

END OF QUALITY CONTROL REPORT

LABORATORY REPORT

This report shall only be reproduced in full with written consent from Clean Water Lab, Inc.